



Declaration of Conformity

Manufacturer:	Authorized Representative:	Notified Body:
ResMed Pty. Ltd. 1 Elizabeth Macarthur Drive Bella Vista NSW 2153 Australia	ResMed SAS Parc Technologique de Lyon 292 Allée Jacques Monod 69791 Saint Priest Cedex France	TÜV SÜD Product Service GmbH Ridlerstraße 65 80339 München Germany

Product: AirTouch N30i

Intended Use:
The AirTouch N30i mask is intended for patients weighing more than 30 kg, who have been prescribed non-invasive CPAP or bi-level positive airway pressure (PAP) therapy. The mask is intended for single-patient reuse in the home environment and multi-patient reuse in the hospital/institutional environment.

Classification: IIa according to Rule 2

EMDN: R0301010201 CPAP Masks

Conformity Assessment Route: Annex IX (excluding Chapter II), Regulation EU 2017/745

Basic UDI-DI: 619498EC1816P

Common Specification: N/A

We herewith declare that the above mentioned products are in conformity with the Council Regulation 2017/745 for medical devices.

Compliance is applicable from the date listed below. All supporting documentation is retained at the premises of the manufacturer. This declaration is issued under the sole responsibility of ResMed Pty. Ltd.

EC Certificate Number: G15 049861 0261 Rev. 01

SRN: AU-MF-000011753

Signed at Sydney, Australia on: 18 February 2026

DocuSigned by:
Nicole Wilson
F9F55744DEA64A0...

Nicole Wilson
Person Responsible for Regulatory Compliance (PRRC)
ResMed Pty. Ltd.

EC181c
First issued: 09 September 2024